



Material Safety Data Sheet (MSDS)

Boston Synthetic Vinyl Exam Gloves

02/2005

(Powder Free) Item Numbers 2511, 2512, 2513, 2514

Physical Data - Composition

Polyvinyl Chloride	48.54%
Di-Isononyl Phthalate (DINP)	41.26%
Di-Octyl Adipate (DOA)	4.85%
Pentanediol Diisobutyrate (TXIB)	2.43%
Epoxidized Soy Bean Oil	1.46%
Calcium-Zinc Stabilizer	1.46%

Fire & Explosion

Stability: Stable

Flash Point: N/A

Extinguishing Media: Foam, Carbon Dioxide, Water

Precaution

This product contains polyvinyl chloride (PVC). If the user or a patient is allergic to vinyl or PVC or experience any discomfort, discontinue use immediately and consult with a physician.

Do not reuse gloves. Good quality exam gloves provide an excellent biological barrier. They are not intended for use as a chemical barrier.

Storage

Store this product in original packing in a cool, dry, and well-ventilated area, away from dust, sunlight, moisture, X-ray, and excessive heat above 100°F (37°C).

(Powdered) Item Numbers 2211, 2212, 2213, 2214

Physical Data - Composition

Polyvinyl Chloride	48.54%
Di-Isononyl Phthalate (DINP)	41.26%
Di-Octyl Adipate (DOA)	4.85%
Pentanediol Diisobutyrate (TXIB)	2.43%
Epoxidized Soy Bean Oil	1.46%
Calcium-Zinc Stabilizer	1.46%

Donning Powder: Absorbable Dusting Powder (USP) Generic Name- Modified Corn Starch, Magnesium Oxide.

Fire & Explosion

Stability: Stable

Flash Point: N/A

Extinguishing Media: Foam, Carbon Dioxide, Water

Precaution

This product contains polyvinyl chloride (PVC). If the user or a patient is allergic to vinyl or PVC or experience any discomfort, discontinue use immediately and consult with a physician.

Do not reuse gloves. Good quality exam gloves provide an excellent biological barrier. They are not intended for use as a chemical barrier.

Storage

Store in original packing in a cool, dry and well ventilated area, away from dust, sunlight, moisture, X-ray, and excessive heat above 100°F (37°C).